



ZOOM® View™ DISK

Selected Episodes Report

Device
Tachy ModeTELIGEN
Monitor + Therapy

Event V-513: 28 May 2012 00:22

VT-1 Event Onset

Avg A Rate: 58 bpm

Avg V Rate: 143 bpm

Detection: Rhythm ID

At V-Detect

Avg A Rate: 65 bpm

Avg V Rate: 147 bpm

Rate Zone: VT-1

Stability: 14 ms

V>A Rate: True

AFib: False

RhythmID Correlated: False

SRD Met: (False, Off)

ATP Timeout: False

Attempt 1, Burst V ATP

Elapsed Time: 00:00:06

ATP Information:

Number of Bursts: 2

Event Ended 00:00:28







