



ZOOM® View™ DISK

**Selected Episodes Report
Induction 1**

Device

TELIGEN

Tachy Mode

Monitor + Therapy

Event V-1: 31 Oct 2008 12:06

Event was Induced

VF Event Onset

Avg A Rate: N/R bpm

Avg V Rate: 314 bpm

Detection: Rate Only

At V-Detect

Avg A Rate: 42 bpm

Avg V Rate: 324 bpm

Rate Zone: VF

Stability: (6 ms, Off)

V>A Rate: (True, Off)

AFib: (False, Off)

RhythmID Correlated: False

SRD Met: (False, Off)

ATP Timeout: False

Attempt 1, 31 J V Shock

Elapsed Time: 00:00:02

Shock Information:

Charge Time: 5.6 s

Lead Impedance: 40 Ω

Lead Polarity: Initial

Event Ended 00:00:40



