

**Event V-1: 27 Mar 2012 10:08**

Event was Induced

### VF Event Onset

Avg V Rate: 303 bpm

Detection: Rate Only

Template: N/R

### At V-Detect

Avg V Rate: 420 bpm

Rate Zone: VF

RhythmID Correlated: False

RhythmMatch™: N/R

SRD Met: (False, Off)

ATP Timeout: False

### Attempt 1, 21 J V Shock

Elapsed Time: 00:00:03

### Aborted Attempt

No shock attempted due to fail to reconfirm

### At V-Detect

Avg V Rate: 294 bpm

Rate Zone: VF

ATP Timeout: False

### Attempt 2, 21 J V Shock

Elapsed Time: 00:00:12


### Shock Information:


Charge Time: 0.2 s

Lead Impedance: 53  $\Omega$


Lead Polarity: Initial

**Event Ended 00:00:44**





VF 145 VF 138 VS 305 VF 143 VS 325 VF 168 VF 270 VS 290 VF 143 VF 150 VF 270 VF 150 VF 255 -- VS 293 VS 305 VS 350 VF 195 VF 198 VF 235 VS 278 VS 288 VS 305 VF 258 VF 208 VF 185  
3 VF 138 VF 145 VF 158 VF 138 VF 143 VF 140 VF 153 VF 148  
V-Detect Chrg Dvrt



VF 138 VS 275 VS 358 VF 138 VF 160 VS 378 VF 143 VS 278 VF 180 VF 253 -- VS 753 VS 780 VS 780 VS 748  
8 VF 155 VF 138 VF 145 VF 153 -- VF 140  
V-Detect Chrg Chrg  
21J Shk



\_\_\_\_\_

