

## Episode: SVT (155 bpm / 385 ms)

VT/VF Episode 1 of 1

Page 1 of 2

Duration Detection 00:08 (M:S) Alerts SVT discriminators disagree  
Criteria 150 - 213 bpm



### Therapy

### Results

No therapies were delivered

### Diagnosis Summary

	Diagnosis
	<b>SVT</b>
Time to Diagnosis	5.75 sec
Rate (CL)	155 bpm (385 ms)
Zone	VT-1
VT Diagnosis Criteria	All
Rate Branch Classification	AF/AFL (V < A) Rate Branch

**Morphology** On, ≥ 50% is a match, ≥ 5 matches indicate SVT  
Min Match Score 51%  
Max Non-Match Score n/a  
No. Template Matches 8 of 8 **(SVT Indicated)**

**Interval Stability** On w/AVA, ≥ 45 ms or AVA < 60 ms indicates SVT  
Stability Delta 20 ms  
AV Association Delta: 130 ms **(VT Indicated)**

### SVT Criteria Statistics

SVT Diagnoses during this episode 1

### Measured SVT Discriminator values for Episode

Min Match Score	51 %	Min/Max Stability Delta	20 ms / 20 ms
Max Non-Match Score	0 %	Min/Max AV Association Delta	130 ms / 130 ms
Template Used	31 Mar 2010 16:21		

### DeFT Response™ Settings (Shock Waveform)

Waveform	Biphasic	VF Shocks (Defib)	1st Phase	2nd Phase
Waveform Mode	Tilt	VT Shocks (CVRT)	65 %	65 %
Shock Configuration	RV to Can		Same as Defib	Same as Defib
RV Polarity	Anode(+)			
Defib Max Sensitivity	0.3 mV			

# Episode: SVT (155 bpm / 385 ms)

VT/VF Episode 1 of 1

Page 2 of 2

1: A Sense Amp AutoGain (1.4 mm/mV)

3: Markers

2: V Sense Amp AutoGain (0.4 mm/mV)

Sweep Speed: 25 mm/s

