

Tachy Episode (Archive)

Tachy Episode 4 of 6		Type: VT (Therapy Delivered)	Episode Duration: 0:08
Therapy:	Results:	Alerts: 1	
ATP	Below Rate Detection (CL 640 ms)	≥3 VT/VF episodes in 24 hours	
ATP Therapy Details: VT			
Successful BCL: 232 ms Burst 1: 232, 232, 232, 232, 232, 232, 232 ms			

<p style="text-align: center;">Diagnosis: VT</p> <p style="text-align: center;">Time to Diagnosis: 3.25 sec CL: 290 ms/206 bpm</p> <p>VT Diagnosis Criteria: 2 of 3</p> <p>Morphology:</p> <p>Programmed: On, ≥ 60% is a match, ≥ 5 of 8 matches indicate SVT Template Updated: 12 Oct 2006 23:55</p> <p>Measured:</p> <p>Min Match Score: N/A</p> <p>Max Non-Match Score: 0%</p> <p>No. Template Matches: 0 of 8 (VT Indicated)</p> <p>Interval Stability:</p> <p>Programmed: On, ≥ 50 ms indicates SVT (SIH Passive, ≥ 2 intervals indicates SVT)</p> <p>Measured:</p> <p>Stability Delta: 10 ms (VT Indicated)</p> <p>SIH Count: 0 (Passive: VT Indicated)</p> <p>Sudden Onset:</p> <p>Programmed: On, <80 ms indicates SVT</p> <p>Measured Max Delta: 400 ms (VT Indicated)</p>	
<p style="text-align: center;">SVT Criteria Statistics</p> <p>Number of SVT Diagnoses during this Episode: 0</p> <p>Measured SVT Discriminator values for Episode:</p> <p>Min Match Score: N/A</p> <p>Max Non-Match Score: 0%</p> <p>Min / Max Stability Delta: 10 ms / 10 ms</p> <p>Min / Max SIH Count: 0 / 0</p> <p>Max Onset Delta: 400 ms</p>	

Atlas™ VR

Tachy Episode (Archive)

EGM 1: V Sense / Pace (1.4 mm/mV, ±8.9 mV)

EGM 2: Off

Sweep Speed: 25.0 mm/s

