



Episode: VF (315 bpm / 190 ms)

VT/VF Episode 23 of 36

Page 1 of 3

Duration 00:12 (M:S)
Detection Criteria 222 bpm

Alerts



Non-sustained episode
Three or more VT/VF episodes in 24 hours

Therapy

1) Aborted Defib
36.0 J (830V)

Results

1) Below Rate Detection
(CL 690 ms)

Diagnosis Summary

Diagnosis

VF
Time to Diagnosis 2.25 sec
Rate (CL) 315 bpm (190 ms)

DeFT Response™ Settings (Shock Waveform)

Waveform Biphasic
Waveform Mode Tilt
Shock Configuration RV to Can
RV Polarity Anode(+)
Defib Max Sensitivity 0.5 mV

VF Shocks (Defib)
VT Shocks (CVRT)

1st Phase

65 %
Same as Defib

2nd Phase

65 %
Same as Defib

Episode: VF (315 bpm / 190 ms)

VT/VF Episode 23 of 36




Page 2 of 3

1: A Sense Amp AutoGain (1.3 mm/mV)

3: Markers

2: V Sense Amp AutoGain (0.4 mm/mV)

Sweep Speed: 25 mm/s



Episode: VF (315 bpm / 190 ms) (Continued)

VT/VF Episode 23 of 36

Page 3 of 3

1: A Sense Amp AutoGain (1.3 mm/mV)

3: Markers

2: V Sense Amp AutoGain (0.4 mm/mV)

Sweep Speed: 25 mm/s

