

Episode: VF (279 bpm / 215 ms)

VT/VF Episode 1 of 1

Page 1 of 3

Duration 00:18 (M:S) No Alerts
Detection Criteria 200 bpm

Therapy	Results	HV Therapy	
1) Defib 36.0 J (844V)	1) Below Rate Detection (CL > 995 ms)	Last HV Lead Impedance	43 Ω
		Last Charge Time	7.5 sec
		Delivered PW	+5.0 ms, -5.0 ms

Diagnosis Summary

Diagnosis

VF
Time to Diagnosis 3.75 sec
Rate (CL) 279 bpm (215 ms)

DeFT Response™ Settings (Shock Waveform)			1st Phase	2nd Phase
Waveform	Biphasic	VF Shocks (Defib)	65 %	65 %
Waveform Mode	Tilt			
Shock Configuration	RV to SVC & Can			
RV Polarity	Anode(+)			
Defib Max Sensitivity	0.5 mV			

Episode: VF (279 bpm / 215 ms)

VT/VF Episode 1 of 1

Page 2 of 3

1: A Sense Amp AutoGain (0.8 mm/mV)

3: Markers

2: V Sense Amp AutoGain (0.4 mm/mV)

Sweep Speed: 25 mm/s



