


Episode: SVT (142 bpm / 420 ms)

VT/VF Episode 1 of 3

Page 1 of 4

Duration Detection Criteria	01:38 (M:S) 141 - 229 bpm	Alerts 	Duration over 1 minute
-----------------------------	------------------------------	---------------------------------------------------------------------------------------------	------------------------

Therapy	Results
No therapies were delivered	

Diagnosis Summary

	Diagnosis
	SVT
Time to Diagnosis	9.25 sec
Rate (CL)	142 bpm (420 ms)
Zone	VT-1 (Monitor)
VT Diagnosis Criteria	All

Sudden Onset	On, < 22 % indicates SVT
Max Delta	2 % (SVT Indicated)

Interval Stability	On w/SlH, ≥ 45 ms or SlH Count ≥ 2 intervals indicates SVT
Stability Delta	235 ms (SVT Indicated)

SVT Criteria Statistics

SVT Diagnoses during this episode 17

Measured SVT Discriminator values for Episode

Min/Max Stability Delta	190 ms / 240 ms	Max Onset Delta	2 %
-------------------------	-----------------	-----------------	-----

DeFT Response™ Settings (Shock Waveform)

Waveform	Biphasic	VF Shocks (Defib)	1st Phase	2nd Phase
Waveform Mode	Tilt	VT Shocks (CVRT)	65 %	65 %
Shock Configuration	RV to Can		Same as Defib	Same as Defib
RV Polarity	Anode(+)			
Defib Max Sensitivity	0.3 mV			

Episode: SVT (142 bpm / 420 ms)




VT/VF Episode 1 of 3

Page 2 of 4

1: V Sense Amp AutoGain (0.7 mm/mV)

2: Markers

Sweep Speed: 25 mm/s



Episode: SVT (142 bpm / 420 ms) (Continued)



VT/VF Episode 1 of 3

Page 3 of 4

1: V Sense Amp AutoGain (0.7 mm/mV)

2: Markers

Sweep Speed: 25 mm/s



Episode: SVT (142 bpm / 420 ms) (Continued)

VT/VF Episode 1 of 3

Page 4 of 4

1: V Sense Amp AutoGain (0.7 mm/mV)

2: Markers

Sweep Speed: 25 mm/s

