

Device: **Reveal® DX 9528**

Date of Interrogation:

Parameters

	Detection	ECG Recording	Interval (Rate)	Duration
FVT	On	On	260 ms (231 bpm)	30/40 beats
VT	On	On	340 ms (176 bpm)	16 beats
Asystole	On	On		3 sec
Brady	On	On	2000 ms (30 bpm)	4 beats

Sensing

Sensitivity	0.035 mV (35 µV)
Blank after Sense	70 ms
Sensing Threshold Decay Delay	100 ms

Detection Enhancements

VT Detection	
VT Stability	Off
VT Onset	Off

Device Data Collection

Device Data Collection	On
Device Date/Time	10-Oct-2010 19:15

Device Information

Device	Medtronic	REVEAL DX 9528	Implanted:
--------	-----------	----------------	------------

Notes

syncope

Device: **Reveal® DX 9528**

Date of Interrogation:

	Prior to Last Session 16-Apr-2010 to 11-Jul-2010 86 days	Since Last Session 11-Jul-2010 to 10-Oct-2010 91 days	Device Lifetime Total (Since 16-Apr-2010) 6 months
Episode Counters			
Symptom	0	0	0
FVT	0	0	0
VT	0	0	0
Asystole	0	0	0
Brady	0	0	0