


SVT/NST Episode #66 Report

ICD Model: InSync Sentry 7298

Serial Number:

Date of Visit: Dec 07, 2009

ID#	Date/Time	A. Cycle	V. Cycle	Duration	Reason
66	Jun 19 19:03:25	180 ms	360 ms	49 sec	AFib/AFlutter



SVT/NST Episode #66 Report

ICD Model: InSync Sentry 7298

Serial Number:

Date of Visit: Dec 07, 2009

Episode #66 - Jun 19, 2009 19:03:25

Episode Summary

A. Median Cycle 180 ms
V. Median Cycle 360 ms
V. Average Cycle 360 ms
SVT Duration 49 sec

Device was in Mode Switch during SVT/NST episode.

SVT Criteria Triggered

AFib/AFlutter

Parameter Settings

	Enable	Initial	Redetect	V Interval (Rate)
VF	On	12/16	9/12	300 ms (200 bpm)
FVT	via VF			250 ms (240 bpm)
VT	On	16	12	370 ms (162 bpm)

PR Logic

AFib/AFlutter On
Sinus Tach On
Other 1:1 SVTs On
SVT Limit 300 ms

3

Other Enhancements

Stability 40 ms

Additional Settings

1:1 VT-ST Boundary 50 %
High Rate Timeout Off

Sensitivity

Atrial 0.3 mV
Ventricular 0.3 mV

	EGM 1	EGM 2
EGM Source	Atip to Aring	RVtip to RVring
EGM Range	+/- 8 mV	+/- 8 mV