



ZOOM® View™ DISK

**Selected Episodes Report  
Therapie 4**Device COGNIS  
Tachy Mode Monitor + Therapy**Event V-67: 10 Dec 2011 23:53****VT Event Onset**Avg A Rate: 80 bpm  
Avg V Rate: 171 bpm  
Detection: Rate Only**At V-Detect**Avg A Rate: 86 bpm  
Avg V Rate: 201 bpm  
Rate Zone: VT  
Stability: (33 ms, Off)  
V>A Rate: (True, Off)  
AFib: (False, Off)  
SRD Met: (False, Off)  
ATP Timeout: False  
Onset Intvl: (170 ms, Off)  
Onset %: (24 %, Off)**Attempt 1, Non Therapy V**

Elapsed Time: 00:00:05

**Attempt Information:**


Tachy therapy is programmed off.


**At V-Detect**Avg A Rate: 85 bpm  
Avg V Rate: 240 bpm  
Rate Zone: VF  
ATP Timeout: False**Attempt 2, 41 J V Shock**

Elapsed Time: 00:00:12


Aborted Attempt

VF ATP delivered, no shock attempted due to fail to reconfirm


**Event Ended 00:00:26**



AS 710		AS 708		AS 703		AS 680		(AS) 613		(AS) 620		(AS) 623		(AS) 623		AS 625		AS 628		AS 635		AS 765		
VT 313	RVS 378	VT 345	VF 280	VF 255	VT 308	VT 303	VT 313	VT 290	VT 290	VT 293	VT 315	VT 300	VF 283	VT 288	VT 308	VT 295	VT 358	VT 315	VT 303	VF 280	VF 265	VF 258	VF 253	VF 21
LVS 28	LVS 35	LVS 33	LVS 30	LVS 30	LVS 35	LVS 33	LVS 33	LVS 35	LVS 35	LVS 30	LVS 33	LVS 30	LVS 33	LVS 30	LVS 35	LVS 33	LVS 35	LVS 30	LVS 33	LVS 30	LVS 30	LVS 30	LVS 30	LVS 3
V-Epsd								V-Detect PVP→		PVP→		PVP→		PVP→						V-Dur				



AS 723		AS 700		AS 683		AS 673		AS 660		--		AS 615		AS 603		AS 590		AS 600		AS 603		AS 600	
VF 275	VF 268	VF 248	VF 243	VF 245	VF 260	VF 253	RVP 220	RVP 220	RVP 220	RVP 220	RVP 220	RVP 220	RVS 605	RVS 605	RVS 605	RVP 568	RVP 600	RVP 603	RVP 600	RVP 603	RVP 603	RVP 600	RVP 600
LVS 30	LVS 30	LVS 30	LVS 28	LVS 30	LVS 30	LVS 30	LVS 30	LVS 30	LVS 30	LVS 30	LVS 30	LVS 30	LVS 45	LVS 55	LVS 55	LVP 0	LVP 0	LVP 0	LVP 0	LVP 0	LVP 0	LVP 0	LVP 0
V-Detect																							
Dvrt PVP→																							



AS 615	AS 628	AS 790	AS 705	AS 710	AS 715	AS 705	AS 685	AS 680	AS 688	AS 67
RVP 608 LVP 0	RVP 615 LVP 0	RVP 628 LVP 0	RVP 790 LVP 0	RVP 705 LVP 0	RVP 710 LVP 0	RVP 715 LVP 0	RVP 705 LVP 0	RVP 685 LVP 0	RVP 680 LVP 0	RVP 688 LVP 0

